

## PATENT COOPERATION TREATY

PCT

## INTERNATIONAL PRELIMINARY EXAMINATION REPORT

REC'D 01 APR 2004

WIPO PCT

(PCT Article 36 and Rule 70)

Applicant's or agent's file reference 11882.0024.P	FOR FURTHER ACTION See Notification of Transmittal of International Preliminary Examination Report (Form PCT/IPEA/416)	
International application No. PCT/US03/20268	International filing date (day/month/year) 26 June 2003 (26.06.2003)	Priority date (day/month/year) 26 June 2002 (26.06.2002)
International Patent Classification (IPC) or national classification and IPC IPC(7): A61M 1/12 and US Cl.: 600/17		
Applicant MICROMED TECHNOLOGY, INC.		

1. This international preliminary examination report has been prepared by this International Preliminary Examining Authority and is transmitted to the applicant according to Article 36.
2. This REPORT consists of a total of <u>1</u> sheets, including this cover sheet.
<input type="checkbox"/> This report is also accompanied by ANNEXES, i.e., sheets of the description, claims and/or drawings which have been amended and are the basis for this report and/or sheets containing rectifications made before this Authority (see Rule 70.16 and Section 607 of the Administrative Instructions under the PCT). These annexes consist of a total of <u>1</u> sheets.
3. This report contains indications relating to the following items:
<ul style="list-style-type: none"> <li>I <input checked="" type="checkbox"/> Basis of the report</li> <li>II <input type="checkbox"/> Priority</li> <li>III <input type="checkbox"/> Non-establishment of report with regard to novelty, inventive step and industrial applicability</li> <li>IV <input checked="" type="checkbox"/> Lack of unity of invention</li> <li>V <input checked="" type="checkbox"/> Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement</li> <li>VI <input type="checkbox"/> Certain documents cited</li> <li>VII <input type="checkbox"/> Certain defects in the international application</li> <li>VIII <input type="checkbox"/> Certain observations on the international application</li> </ul>

Date of submission of the demand 20 January 2004 (20.01.2004)	Date of completion of this report 23 February 2004 (23.02.2004)
Name and mailing address of the IPEA/US Mail Stop PCT, Attn: IPEA/US Commissioner for Patents P.O. Box 1450 Alexandria, Virginia 22313-1450 Facsimile No. (703)305-3230	Authorized officer Jeffrey R. Jastrzab Telephone No. (703) 308-0858

## I. Basis of the report

## 1. With regard to the elements of the international application:\*

the international application as originally filed.

the description:  
pages 1-25 as originally filed  
pages NONE, filed with the demand  
pages NONE, filed with the letter of \_\_\_\_\_

the claims:  
pages 26-29, as originally filed  
pages NONE, as amended (together with any statement) under Article 19  
pages NONE, filed with the demand  
pages NONE, filed with the letter of \_\_\_\_\_

the drawings:  
pages 1-14, as originally filed  
pages NONE, filed with the demand  
pages NONE, filed with the letter of \_\_\_\_\_

the sequence listing part of the description:  
pages NONE, as originally filed  
pages NONE, filed with the demand  
pages NONE, filed with the letter of \_\_\_\_\_

## 2. With regard to the language, all the elements marked above were available or furnished to this Authority in the language in which the international application was filed, unless otherwise indicated under this item. These elements were available or furnished to this Authority in the following language \_\_\_\_\_ which is:

the language of a translation furnished for the purposes of international search (under Rule 23.1(b)).

the language of publication of the international application (under Rule 48.3(b)).

the language of the translation furnished for the purposes of international preliminary examination (under Rules 55.2 and/or 55.3).

## 3. With regard to any nucleotide and/or amino acid sequence disclosed in the international application, the international preliminary examination was carried out on the basis of the sequence listing:

contained in the international application in printed form.

filed together with the international application in computer readable form.

furnished subsequently to this Authority in written form.

furnished subsequently to this Authority in computer readable form.

The statement that the subsequently furnished written sequence listing does not go beyond the disclosure in the international application as filed has been furnished.

The statement that the information recorded in computer readable form is identical to the written sequence listing has been furnished.

4.  The amendments have resulted in the cancellation of:

the description, pages NONE

the claims, Nos. NONE

the drawings, sheets/fig NONE

5.  This report has been established as if (some of) the amendments had not been made, since they have been considered to go beyond the disclosure as filed, as indicated in the Supplemental Box (Rule 70.2(c)).\*\*

\* Replacement sheets which have been furnished to the receiving Office in response to an invitation under Article 14 are referred to in this report as "originally filed" and are not annexed to this report since they do not contain amendments (Rules 70.16 and 70.17).

\*\* Any replacement sheet containing such amendments must be referred to under item 1 and annexed to this report.

## INTERNATIONAL PRELIMINARY EXAMINATION REPORT

International application No.

PCT/US03/20268

## IV. Lack of unity of invention

1. In response to the invitation to restrict or pay additional fees the applicant has:
  - restricted the claims.
  - paid additional fees.
  - paid additional fees under protest.
  - neither restricted nor paid additional fees.
2.  This Authority found that the requirement of unity of invention is not complied with and chose, according to Rule 68.1, not to invite the applicant to restrict or pay additional fees.
3. This Authority considers that the requirement of unity of invention is accordance with Rules 13.1, 13.2 and 13.3 is:
  - complied with.
  - not complied with for the following reasons:

This application contains the following inventions or groups of inventions which are not so linked as to form a single general inventive concept under PCT Rule 13.1. In order for all inventions to be examined, the appropriate additional examination fees must be paid.

Group I, claim(s) 1, 16 and 17 drawn to time and frequency domain flow measurements.

Group II, claim(s) 2, 16 and 19 drawn to measuring instantaneous pump power or current.

Group III, claim(s) 3-6, 10-12 and 16 drawn to resting heart rate and exercise.

Group IV, claim(s) 7-8, 10-12, 15 and 16 drawn to control having constant flow, amplitude and speed..

Group V., claim(s) 9 and 16, drawn to unloading.

Group VI, claim(s) 13 and 14, drawn to diastolic flow.

Group VII, claim(s) 18, drawn to dQ/dn.

The inventions listed as Groups I-VII do not relate to a single general inventive concept under PCT Rule 13.1 because, under PCT Rule 13.2, they lack the same or corresponding special technical features for the following reasons: each of the methods requires different feedback for flow control thus making them independent.

4. Consequently, the following parts of the international application were the subject of international preliminary examination in establishing this report:

- all parts.
- the parts relating to claims Nos. \_\_\_\_

## INTERNATIONAL PRELIMINARY EXAMINATION REPORT

International application No.  
PCT/US03/20268V. Reasoned statement under Rule 66.2(a)(ii) with regard to novelty, inventive step or industrial applicability;  
citations and explanations supporting such statement

## 1. STATEMENT

Novelty (N)

Claims 1-19 YES  
Claims NONE NO

Inventive Step (IS)

Claims 1-19 YES  
Claims NONE NO

Industrial Applicability (IA)

Claims 1-19 YES  
Claims NONE NO

## 2. CITATIONS AND EXPLANATIONS

Claims 1-19 meet the criteria set out in PCT Article 33(2)-(3), because the prior art does not teach or fairly suggest controlling a blood pump using instantaneous flow data in both the time and frequency domain. Note that "X,E" references are not applicable in the Preliminary Examination report.

Claims 1-19 meet the criteria set out in PCT Article 33(4), and thus have industrial applicability because the subject matter claimed can be made or used in industry.

----- NEW CITATIONS -----